

MAR 28 2006

510(k) Summary for N Antisera to Human α_1 -Antitrypsin

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) number is: K05 3072

1. Manufacturer's Name, Address, Telephone, and Contact Person, Date of Preparation:

Manufacturer: Dade Behring Marburg GmbH
Emil-von-Behring Str. 76
D-35001
Marburg, Germany

Contact Information: Dade Behring Inc.
Glasgow Site
P.O. Box 6101
Newark, Delaware 19714
Attn: Kathleen Dray-Lyons
Tel: 781-826-4551
Fax: 781-826-2497

Preparation date: October 28, 2005

- 2. Device Name/ Classification:** N Antisera to Human α_1 -Antitrypsin
Class: α_1 -Antitrypsin Immunological Test System, Class II,
21 CFR 866.5130
Panel: Immunology
Product Code: DEM

3. Identification of the Legally Marketed Device:

N Antisera to Human α_1 -Antitrypsin– K860894

4. Device Description:

Proteins contained in human body fluids form immune complexes in an immunochemical reaction with specific antibodies. These complexes scatter a beam of light passed through the sample. The intensity of the scattered light is proportional to the concentration of the relevant protein in the sample. The result is evaluated by comparison with a standard of known concentration.

000036

5. Device Intended Use:

In vitro diagnostic reagents for the quantitative determination of α_1 -antitrypsin in human serum, heparinized and EDTA plasma by means of immunonephelometry on the BN™ Systems.

6. Medical device to which equivalence is claimed and comparison information:

The modified N Antisera to Human α_1 -Antitrypsin assay is substantially equivalent to the N Antisera to Human α_1 -Antitrypsin currently marketed (K860894). The modified N Antisera to Human α_1 -Antitrypsin assay, like the current N Antisera to Human α_1 -Antitrypsin assay, is intended for the quantitative determination of α_1 -antitrypsin by means of immunonephelometry on the BN™ Systems.

7. Device Performance Characteristics:

To demonstrate equivalence in measurement between serum and heparinized or EDTA plasma, method comparisons were performed. The studies demonstrate equivalent performance with correlation coefficients between 0.97 and 0.99



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

MAR 28 2006

Dade Behring, Inc.
c/o Ms. Kathleen A Dray-Lyons
Glasgow Site
P.O. Box 6101
Newark, DE 19714

Re: k053072

Trade/Device Name: N Antisera to Human α 1-Antitrypsin Assay
Regulation Number: 21 CFR 866.5130
Regulation Name: Alpha-1-Antitrypsin Immunological Test System
Regulatory Class: Class II
Product Code: DEM
Dated: October 28, 2005
Received: November 1, 2005

Dear Ms. Dray-Lyons:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 –

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0484. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Robert L. Becker, Jr.", with a stylized flourish at the end.

Robert L. Becker, Jr., M.D., Ph.D.

Director

Division of Immunology and Hematology Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

Center for Devices and Radiological Health

Enclosure

Indications Statement

K053072

Device Name: N Antisera to Human α_1 -Antitrypsin Assay

Indications for Use:

In vitro diagnostic reagents for the quantitative determination of α_1 -antitrypsin in human serum, heparinized and EDTA plasma by means of immunonephelometry on the BN™ Systems. The measurement of α_1 -antitrypsin aids in the diagnosis of several conditions including adult cirrhosis of the liver. In addition, α_1 -antitrypsin deficiency has been associated with pulmonary emphysema in conjunction with other laboratory and clinical findings.

Prescription Use X
(Per 21 CFR 801 Subpart D)

Over-The-Counter-Use _____
(21 CFR 801)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Maria Chan
Division Sign-Off

Page 1 of ____

Office of In Vitro Diagnostic Device
Evaluation and Safety

053072